



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUN 23 1998

Mr. Keith A. Roberts
Porous Media Corporation
1350 Hammond Road
St. Paul, MN 55110

Re: K982127
Porous Media DEF25™ Disposable Respiratory Filter
Regulatory Class: II (two)
Product Code: 73 CAH
Dated: June 16, 1998
Received: June 17, 1998

Dear Mr. Roberts:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 - Mr. Keith A. Roberts

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4648. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597, or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,

A handwritten signature in black ink, reading "Thomas J. Callahan". The signature is fluid and cursive, with the first name "Thomas" being the most prominent part.

Thomas J. Callahan, Ph.D.
Director
Division of Cardiovascular,
Respiratory, and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

K982127

Page 1 of 1

510(k) Number (if known): K964979 (Previously Cleared Device)

Device Name: DEF25 Disposable Respiratory Filter

Indications For Use:

CONTENTS:

One Disposable Breathing Circuit Bacterial Filter

CAUTION:

FEDERAL (U.S.A.) LAW RESTRICTS THIS DEVICE TO SALE BY OR ON ORDER OF A PHYSICIAN.

INDICATIONS:

- THIS DISPOSABLE FILTER HAS BEEN DESIGNED FOR SINGLE USE AS A MAIN FLOW FILTER, INSPIRATORY AND EXHALATION FILTER. THIS FILTER IS HYDROPHOBIC.
- CLEAR PLASTIC SHELL ALLOWS FOR VISUAL INSPECTION OF THE FILTER MEDIUM.
- FOR USE WITH RESPIRATORY THERAPY AND ANESTHESIA APPLICATIONS.
- HELPS PROTECT PATIENTS FROM RESPIRATORY AND ANESTHESIA VENTILATOR CONTAMINANTS.
- VERSATILE, DISPOSABLE, ECONOMY OF USE - HELPS PROTECT PATIENT, VENTILATORY EQUIPMENT, AND HOSPITAL STAFF.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ☒
 (Per 21 CFR 801.1091)

or Over-The-Counter Use ☐

(Optional Format 1-2-96)

Mark Kramer

(Division Sign-Off)

Division of Cardiovascular, Respiratory,
and Neurological Devices

510(k) Number

K982127